


Code 2934

QAC Dual Range  **50 Test Strips**

Insta-Test[®] ANALYTIC



Using a cup-size sample, immerse strip for **2 seconds**. **SHAKE ONCE TO REMOVE EXCESS WATER**. Wait **15 seconds** and **immediately compare to color chart**.



- Keep wet fingers out of the bottle.
- Cap bottle tightly.
- Store in a cool, dry place.



LaMotte Company
Chester town, MD 21620
<http://www.lamotte.com>

