



Code 2996

Nitrite & Nitrate
50 Test Strips



Insta-Test[®] ANALYTIC

Top pad:	Nitrite as NO ₂ -N ppm	0	0.5	1	5	10
Next pad:	Nitrate as NO ₃ -N ppm	0	5	10	25	50

Using a cup-size sample, immerse strip for **2 seconds** and remove with **pads face up**. **DO NOT SHAKE OFF EXCESS WATER**. Wait **60 seconds** and **immediately compare** to color chart.



- ! Keep wet fingers out of the bottle.
- ! Cap bottle tightly.
- ! Store in a cool, dry place.



LaMotte Company
802 Washington Ave., Chestertown, MD 21620 USA
410-778-3100 • www.lamotte.com

