



# DRINKING WATER 6

Nitrate, Nitrite, pH, Free Chlorine, Total Chlorine & Total Hardness

25 Test Strips

Code 2933-G

## Insta-Test<sup>®</sup> ANALYTIC

Immerse strip for 2 seconds.  
Remove with pads face up.  
**SHAKE ONCE TO REMOVE  
EXCESS WATER.** Immediately  
read **Free Chlorine, Total  
Chlorine, Total Hardness** and  
**pH.** Wait 30 seconds, then  
read: **Nitrite > Nitrate.**

- Keep wet fingers out of the vial
- Close vial tightly after removing strip
- Store in a cool, dry place

LaMotte Company  
Chestertown, Maryland 21620 USA  
[www.lamotte.com](http://www.lamotte.com)



|                                                |  |   |     |     |     |     |     |     |       |             |
|------------------------------------------------|--|---|-----|-----|-----|-----|-----|-----|-------|-------------|
| <b>Free Chlorine</b><br>ppm                    |  | 0 | 0.5 |     | 1   |     | 3   | 5   | 10    |             |
| <b>Total Chlorine</b><br>ppm                   |  | 0 | 0.5 |     | 1   |     | 3   | 5   | 10    |             |
| <b>Total Hardness</b><br>ppm (gpg)             |  | 0 | 50  | 100 | 200 | 400 | (0) | (3) | (5.8) | (11.7) (23) |
| <b>pH</b>                                      |  | 4 | 5   | 6   | 7   | 8   | 9   | 10  |       |             |
| <b>Nitrite</b><br>as NO <sub>2</sub> -N<br>ppm |  | 0 | 0.5 | 1   | 5   | 10  |     |     |       |             |
| <b>Nitrate</b><br>as NO <sub>3</sub> -N<br>ppm |  | 0 | 5   | 10  | 25  | 50  |     |     |       |             |