

AMMONIA NITROGEN KIT

Code 4795-01 | Octa-Slide Method, 1-8 ppm



| QUANTITY | CONTENTS | CODE |
|----------|------------------------------------------------|-----------|
| 30 mL | Ammonia Nitrogen Reagent #1 | 4797WT-G |
| 30 mL | *Ammonia Nitrogen Reagent #2 | *4798WT-G |
| 2 | Test Tubes, 2.5-10 mL, plastic, w/caps | 0106 |
| 1 | Octa-Slide 2 Viewer | 1101 |
| 1 | Ammonia Nitrogen Octa-Slide 2 Bar, 1.0-8.0 ppm | 4796-01 |

*Reagent is a potential health hazard. **READ SDS:** lamotte.com
Emergency information:
 Chem-Tel USA 1-800-255-3924
 Int'l, call collect, 813-248-0585




To order individual reagents or test kit components, use the specified code number.



Warning! This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully. Not to be used by children except under adult supervision.

USE OF THE OCTA-SLIDE 2 VIEWER

PROCEDURE

| | | |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>1. Insert Ammonia Nitrogen Octa-Slide 2 Bar [4796-01] into then Octa-Slide 2 Viewer [1101].</p>  | <p>2. Fill test tube [0106] to 5 mL line with sample water.</p>  | <p>3. Add 4 drops of Ammonia Nitrogen Reagent #1 [4797WT]. NOTE: When testing salt water, add 8 drops of Ammonia Nitrogen Reagent #1.</p>  |
| <p>4. Cap and mix.</p>  | <p>5. Add 8 drops of Ammonia Nitrogen Reagent #2 [4798WT].</p>  | <p>6. Cap and mix.</p>  |
| <p>7. Insert test tube into the Octa-Slide 2 Viewer [1101]. Hold the Viewer so that non-direct light enters through the back.</p>  | <p>8. Match sample color to a color standard. Record as ppm Ammonia Nitrogen.</p>  | |